

	Name of Records/Documents	Rule Cross Reference	Time Interval for Keeping Records/Documents
(A)	Current 25 TAC §289.203, 204, 205, 226, 227, and 231	As listed on certificate of registration	Until termination of registration
(B)	Current certificate of registration	§289.203(b)(1)(B)	Until termination of registration
(C)	Notice of violation from last inspection, if applicable	§289.203(b)(1)(D)	Until next inspection
(D)	Documentation of correction of any violations	§289.203(b)(1)(D)	Until next inspection
(E)	Personnel monitoring records	§289.231(m)	Until termination of registration
(F)	Surveys (public dose evaluation)	§289.231(o)	Until termination of registration
(G)	FDA variances on equipment	§289.227(h)	Until termination of registration
(H)	Current operating and safety procedures	§289.227(i)(2)	Until termination of registration
(I)	Protective devices annual check	§289.227(i)(4)(B)	3 years
(J)	Credentials of individuals operating radiation machines <b>[-chiropractic, medical, and podiatric facilities]</b>	§289.227(i)(5)	Until individual leaves that facility
(K)	Entrance exposure rate/fluoroscopy	<u>§289.227(m)(3)(D)</u> <b>[§289.227(o)(3)(D)]</b>	3 years

---

	Name of Records/Documents	Rule Cross Reference	Time Interval for Keeping Records/Documents
(L)	CT dose measurements	<u>§289.227(n)(3)(C)</u> <b>[§289.227(p)(3)(C)]</b>	3 years
(M)	CT films resulting from quality control tests	<u>§289.227(n)(4)(B)</u> <b>[§289.227(p)(4)(B)]</b>	3 years
(N)	Equipment performance evaluations and corrections <b>[-chiropractic, medical and podiatric facilities]</b>	<u>§289.227(o)(1),(2), and (3)</u> <b>[§289.227(q)(1),(2), and (3)]</b>	3 years
(O)	Film processing records and corrections	<u>§289.227(p)(2),(3),and (6)</u> <b>[§289.227(r)(2),(3), and (6)]</b>	3 years
(P)	Alternate processing systems records	<u>§289.227(q)</u> <b>[§289.227(s)]</b>	3 years
<u>(Q)</u>	<u>Digital imaging acquisition systems records</u>	<u>§289.227(r)</u>	<u>3 years</u>
<u>(R)</u> <b>[(Q)]</b>	Receipt, transfer, and disposal	<u>§289.226(m)(1)(D)</u> <b>[§289.231]</b>	Until termination of registration
<u>(S)</u> <b>[(R)]</b>	Calibration and intercomparison of system <b>[of instruments]</b> used to ensure compliance with this chapter	<u>§289.227(o)(4)</u> <b>[§289.227(q)(4)]</b> §289.231	3 years

---